

C. Wies, R. Lebert, B. Jäggle, L. Juschkin,

AIXUV GmbH, Steinbachstrasse 15, D-52074 Aachen, Germany

F. Sobel, H. Seitz, L. Aschke; SCHOTT Lithotec AG, Jerusalem Strasse 13, D-98617 Meiningen, Germany

C. Laubis, F. Scholze; Physikalisch-Technische Bundesanstalt, Abbestrasse 2-12, D-10587 Berlin, Germany

W. Biel; IPP, Forschungszentrum Jülich GmbH, D-52425 Jülich, Germany

O. Steffens; S&F NC-Systemtechnik GbR, Plitschard 172, D-52134 Herzogenrath, Germany

Abstract:

AIXUV GmbH and partners have developed a high speed Reflectometer for EUV mask blanks which is fully compliant with the SEMI-standard P38 for EUV-mask blank metrology.

The system has been installed in June 2004 at SCHOTT Lithotec AG.

It features high throughput, high lateral and spectral resolution, high reproducibility and low absolute uncertainty. Using AIXUV's EUV-LAMP and debris mitigation, low cost-of-ownership and high availability is expected. The spectral reflectance of up to 3 mask-blanks per hour can be measured with at least 20 spots each. The system is push button-controlled. Results are stored in CSV file format.

For a spot size of $0.1 \times 1 \text{ mm}^2$, 2000 spectral channels of 1.6 pm bandwidth are recorded from 11.6 nm to 14.8 nm. The reflectance measurement is based on the comparison of the sample under test to two reference mirrors calibrated at the PTB radiometry laboratory at BESSY II. The three reflection spectra are recorded simultaneously.

For each spot more than 10^7 photons are accumulated in about 20 s, providing statistical reproducibility below 0.2 % RMS. The total uncertainty is below 0.5 % absolute.

Wavelength calibration better than 1 pm RMS over the whole spectral range is achieved by reference to NIST published wavelengths of about 100 xenon emission lines. It is consistent with the wavelength of the krypton 3d-5p absorption resonance at 13.5947 nm to better than 2 pm.

The work was funded by the German Bundesministerium für Bildung und Forschung (BMBF).




Clean room compatibility is realized by housing the tool and the use of special chosen liquid cooled components. The system is open at the top and the bottom for laminar flow.

Centroid wavelength of mask = $(\lambda_1 + \lambda_2)/2$


Wavelength of max reflectivity

Peak reflectivity

Full width at half maximum (FWHM)



Polychromatic (parallel)



Polychromatic reflectometry:

The whole EUV-spectrum is reflected by the sample, spectrally dispersed and detected by a multi-channel detector. The reference spectrum is obtained from the parallel measurement of reference mirrors.

Advantages:


- All photons from the source are used
- The set-up is stationary
- With a CCD camera:
 - Good DQE
 - “Side-information” (scatter, geometry)
- The whole spectral reflectance curve is measured simultaneously. Hence, for measuring 2000 spectral channels with less than 1000 pulses
- Measuring system is mechanically stable.
- With reference mirrors any fluctuations do not influence the relative results.

Disadvantages:

- Instead of a diode: CCD camera: evaluation.
- Data transfer from the CCD limits the throughputs.

With the monochromatic approach, the light is dispersed first, one wavelength channel is selected by a slit and the sample is illuminated. The reflected light is detected by a diode.

Monochromatic (sequential)



Parameter:**Spectral resolution****Wavelength calibration: better****Wavelength reproducibility better****Measuring spot size
(Spatial sub-resolution within the 2mm)****Measuring spectral range****Time to measure one spot best quality****Throughput****Angle of incidence reproducibility****absolute accuracy (MADT)****relative reproducibility of reflectance****Contrast of measurement****MBR achieved:****1.6 pm ($\lambda/\Delta\lambda > 3000$)** **± 2 pm ($\lambda/\Delta\lambda > 3000$)** **± 1 pm ($\lambda/\Delta\lambda = 13.000$)** **$< 100 \mu\text{m} \times 1 \text{mm}$** **11.8 to 14.9 nm
(2048 channels)** **≤ 20 s (reflectors)
 ≤ 1 min (absorbers)****3 samples per hour
with 15 spots per sample** **$< 30'' \Rightarrow \Delta\lambda < 0.1 \%$** **$< 0.5 \%$ absolute** **$< 0.2 \%$ RMS** **> 11 bit with standard set-up
 > 12 bit achievable
first allows to measure
reflectances below 0.1% on
absorbers**

AixUV Reflektometer

Init

Hand

Auto

Extern

Alarm
 Msg

Ad
Level

Log-In

Shutdown

F1: Main | F2: Alarms | F3: Diagnostics | F4: Evaluation | F5: Set-Up

Mode

Auto

→ ◆

Ready

Operations

Start
Auto

Pause
Auto

Stop
Auto

Info (AD) Init No Command active ::CES not running

2004/04/28 14:24:27

Grafische Auswertung

ID: J351-5

Pos.: X 0.000 Y 0.000

Zoom

Full

Freeze

Max Reflection	:	0.000
Central WaveLength:	:	0.000
Max WaveLength	:	0.000
Return Value	:	FAILURE

Idle

Calc

Push Button controlled operation

AixUV Reflektometer

running | Hand | Auto | Extern

Alarm | Msg | **Ad** Level

Log-In | Shutdown

F1: Main | F2: Alarms | F3: Diagnostics | F4: Evaluation | F5: Set-Up

PLC
 Run
 Fault
 Check
 State Handling
 T 193

Turbo Pumps
 Load Lock
 Func. Chamb
 On
 Fault
 On/Off

Scroll Pumps
 Load Lock
 Func. Chamb
 On
 Fault
 On/Off

Measure
 mks 901
 Dose Meas.

Camera
 Run
 Check

Load Lock
 Arm retracted
 Arm
 Gate Open
 Gate Closed
 Toggle Gate
 Move Arm

Gas
 Nitro. Opened
 Nitro. Closed
 Argon Opened
 Argon Closed
 Nitrogen Open/Close
 Argon Open/Close

Axes
 XY Table
 Z Drives
 On
 On
 Fault
 Fault

Language / Sprache
 English / Englisch
 German / Deutsch

2004/05/06 08:50:38


Grafische Auswertung

ID: J651-1
 Pos.: X 0.000 Y 0.000


Zoom | Full | Freeze

Max Reflection : 0.000
 Central WaveLength: 0.000
 Max WaveLength : 0.000
 Return Value : FAILURE

Idle
 Calc



Focusing grating allows resolving of inhomogeneity within illuminated area.
 This image was taken at the edge of the coating.



Wavelength reproducibility
 Calibration by emission lines


Accuracy of Absolute Values
 Measurement relative to calibrated mirrors

Adapted spectrograph
 $\lambda/\Delta\lambda > 3000$ with special designed concave holographic grating



Reproducibility and Throughput
 Photon statistics better 0.1 %


Small Measurement spot
 down to $20 \mu\text{m} \times 2 \text{mm}$

Avoiding contamination
 Background laminar gas flow,
 Only sides contacts of blank




Contact with the mask is only at the bevels to avoid contamination and damage of the mask blank.






Measurement of a calibrated sample allows to check calibration of MBR and re-calibration of reference mirrors


The inset shows the range from 12 nm to 15 nm.




Spectral channel of 1.6 pm allows to detect a single line as wavelength standard: HERE: Xenon 13.5072 nm line




Upper limit for accuracy = resolution! (monochromator / spectrograph).
 Design guideline $\lambda/\Delta\lambda \geq \text{precision! } 0.005 \text{ nm} \Rightarrow \lambda/\Delta\lambda \geq 2700 \text{ !!}$
 Achieved: > 3000 !




Wavelength scale checked with Kr 3d-5p resonance absorption line at 13.5947 nm. The wavelength scale was calculated using the emission lines. The position of the absorption line was found with less than 2 pm uncertainty.




Reproducibility of the measurement of central wavelength, wavelength of max. reflectance and max. reflectance with the MBR demonstrated with 8 measurements.




MBR data binned to same point distance is nearly as smooth as reflectance curve determined by PTB at BESSY II.



Absolute accuracy and reproducibility is demonstrated in series of measurements on same sample (2 different sites).



Measurement of low reflecting sample (red line, 40 sec. of illumination) in comparison to PTB measurements at 3 different position on the same sample. The small peak at 13.34 nm was due to a reflex of visible light coming out of the pressure gauge.




Reflectance curve measured with the MBR and compared to the same measurement done by PTB at BESSY II.

	λ_{central}	R_{max} (absolute)
PTB	13,168 nm \pm 0.0019 nm	67.72 % \pm 0.1 %
CXRO	13.1684 nm \pm 0.0019 nm	67.63 % \pm 0.1 %
NIST	13.1763 nm \pm 0.01 nm	67.03 % \pm 0.35 %
ASET	13.1671 nm \pm 0.002 nm	67.42 % \pm 0.17%
HIT	13.16 nm	66.6 %
MBR	\pm 0.0016 nm	\pm 0.5 %

Comparison of the results of measurements of one single mirror at different laboratories. In the table, the precision of the MBR, calculated from measurements of calibrated samples, is indicated.

$$MADT = \bar{X} - X_{\text{Target}}$$

MADT is the Mean Average Deviation from Target



AIXUV 's Reflectometer is 2nd Generation System, designed to fulfill requirements as in SEMI-Standard

- a) One spectral channel of $\lambda/\Delta\lambda > 3000$ (300)
- b) Small spot: 0.1 × 1 mm (2 × 2 mm)
- c) Blank alignment
- d) Fast stages
- e) Maintenance free EUV-source
- f) Engineered as a tool for Semiconductor industry
- g) Fab-compatible Control Architecture
- h) Direct verification at Mask Blank manufacturer (SCHOTT Lithotec)

The MBR is designed for clean room compatibility:

- stainless steel housing
- open at top and bottom for laminar flow
- most components are liquid cooled
- clean room compatible PC, keyboard and touch pad
- exhaust of valves is collected and guided into gray room
- roughing pumps and chiller are placed in gray room